

3rd February, 2021

Minutes

PRE-BID MEETING FOR "SUPPLY, INSTALLATION, TESTING AND COMMISSIONING OF CLOCK LASER & ULTRA-STABLE CAVITY SYSTEM AND FEMTOSECOND OPTICAL FREQUENCY COMB" at IUCAA


The Pre-bid meeting for Supply, Installation, Testing and Commissioning of Clock Laser & Ultra-Stable Cavity System and Femtosecond Optical Frequency Comb was held on 3rd February, 2021 at 1100 hrs in Bhaskara 1. The following members were present:

1. Prof. Subhadeep De
2. Mr. B. V. Sawant

Total three representatives from three different organizations attended the meeting. The questions raised by the agencies were satisfactorily clarified by IUCAA. The queries with response are enclosed in Annexure – I and the list of representatives present as Annexure – II.




B. V. Sawant



Subhadeep De

Committee's reply to pre-bid meeting may be approved.

mt
9.2.21



SONAM RAYCHELI
DIRECTOR

Following points were discussed in Pre-Bid Meeting held on 03.02.2021 at 1100 hrs in Bhaskara – 1.

	Details of the party	Question	Answer
	<p>Ms. Ruchi Daryan Director Simco Global Technology & Systems Ltd. Vasant Kunj, New Delhi-110070 Email ID: rdaryan@simcogroup.in Mb. 9910207575</p>	<p>If the Item 1 & 2 are ordered to different vendors who will perform integration between the items 1 & 2?</p>	<p>System integration of the Items-1 & 2 have to be within the scope of the bidders, even if the items are ordered from different supplier.</p>
		<p>If we are quoting for both the items, do we need to take care of integration?</p>	<p>Yes. Bidders are requested to quote charges for the system integration (if any) separately with in their quote. Finally, IUCAA will have right to choose the system integration if required at the time of placing the purchase order.</p>
		<p>What is the delivery term for this material?</p>	<p>It is understood the items are not off the shelf but customized which may lead to longer delivery time than usual. Hence, the bidder has to quote their required time period for delivery of the goods.</p>
		<p>We have not sold material in India before. Our principal is Germany based and their law does not permit to share the PO copies, whether we can provide you the end user details?</p>	<p>It is requested to share the supporting document for the mentioned law(s). In that case the supplier is requested to provide contact person's name, contact details, and reference purchase order no to were the items were supplied.</p>
		<p>Are you going to provide the Custom Duty Exemption certificate?</p>	<p>Yes, in case, order is placed on Foreign party then only the we will be able to provide Custom Duty Exemption certificate. In case the order is placed on Indian party will be able to provide GST Concession Certificate if the items fall under Published in the Gazette of India, Extraordinary, Part II, Section 3, Sub-Section (i). Notification No. 45/2017-Central Tax (Rate) by Govt. Of India, Ministry of Finance (Department of Revenue) dated on 14th November, 2017.</p>

		<p>Our bank is not ready to give Performance Bank Guarantee when the order is given to foreign OEM.</p>	<p>The order shall mention foreign manufacturer as well as name of the Indian supplier, hence based on that bank guarantee from nationalized bank can be made. IUCAA is routinely following that process.</p>
	<p>Mr. Deepak D Yewale Business Manager Aimil Ltd. Vashi, Navi Mumbai – 400705 Email ID: deepakyewale@aimil.com Mb. 7045943425</p>	<p>Whether L1 will decided separately for the Item 1 and item 2 or combined?</p> <p>If you could mention the range from 100-200MHz, then it will help us to quote.</p>	<p>Two items Item – 1 and 2 are separate hence the final selection will be dealt separately for these two items, based on technical compliance, L1 and compatibility among them.</p> <p>A range between 180-250MHz for the comb line spacing will be acceptable to us.</p>
	<p>Mr. Dipak Kanase Manager- Marketing Western Systems Ghatkopar (W), Mumbai – 400086 Email ID: dipak@wsysind.com Mb. 9004647002</p>	<p>Can we quote Item 1a & 1b separately?</p> <p>Are we exempted for tender fees & EMD under MSME registration?</p> <p>Are you going to give separate PO for item 1a and item 1b?</p>	<p>No, that will be considered as partial quoting and will not comply technically.</p> <p>Yes</p> <p>No, 1a and 1b are just subsets of the entire Item-1, hence separate POs for 1a and 1b is not within the scope.</p>